



COMMONWEALTH of VIRGINIA  
DEPARTMENT OF FORENSIC SCIENCE

CENTRAL LABORATORY  
A Nationally Accredited Laboratory

700 NORTH 5TH STREET  
RICHMOND, VIRGINIA 23219  
(804) 786-6971

CERTIFICATE OF INSTRUMENT ACCURACY

INSTRUMENT: Intox EC/IR II  
SERIAL NUMBER: 010606  
CERTIFICATION DATE: September 13, 2013  
DATE RECERTIFICATION DUE: March 12, 2014

The above listed instrument was found to meet all requirements for accuracy and performance established by the Department of Forensic Science. This certification is valid from the above listed certification date until the above listed date for recertification.

Melissa Kennedy


Program Manager, Calibration & Training

CITY OF RICHMOND  
COMMONWEALTH OF VIRGINIA

The foregoing was acknowledged before me this 23<sup>rd</sup> day of September, 2013 by Melissa Kennedy

  
Ciera N. Boxley, Notary Public

My commission expires: August 31, 2014



TLN  
RJK

**Department of Forensic Science  
Breath Alcohol Section  
Instrument Maintenance History  
1/1/1990 To 9/19/2013**

<b>Instrument Serial Number:</b> 010606	<b>As of 19-Sep-13</b>
---	------------------------

Maintenance Date	Certification Date	Technician Initials	Remarks
13-Sep-13	13-Sep-13	TLN (25387)	MOVED INSTRUMENT TO NEW LOCATION. FIRMWARE/SOFTWARE UPGRADE. CALIBRATED AND CERTIFIED.
07-Jun-13	07-Jun-13	HMS (18910)	CERTIFIED
12-Dec-12	12-Dec-12	HMS (18910)	CERTIFIED, REPLACED DRY GAS STANDARD.
26-Jun-12	26-Jun-12	HMS (18910)	CERTIFIED
19-Jan-12	19-Jan-12	HMS (18910)	CERTIFIED, REPLACED DRY GAS STANDARD.
02-Aug-11	02-Aug-11	HMS (18910)	REPLACED DRY GAS STANDARD, CERTIFIED
08-Feb-11	08-Feb-11	RJP (22428)	CERTIFIED.
22-Oct-09	22-Oct-09	SED (24178)	FIRMWARE/SOFTWARE UPGRADE, CERTIFIED.
12-Aug-09	12-Aug-09	NMJ (22392)	REPLACED TANK TOP LID, CERTIFIED.
10-Dec-08	10-Dec-08	HMS (18910)	CERTIFIED.
12-Nov-08	12-Nov-08	NMJ (22392)	CALIBRATED, CERTIFIED AND ACCEPTED.

TLN  
WK

# INTOX EC/IR II

## Quality Assurance Worksheet

Instrument Serial Number 010606 Worksheet Start Date 9/13/2013

Location Dickenson Co SO

Address 1019 Happy Valley Drive, Clintwood VA, 24228

DFS Technician Timmy Neece, Jr License No. 25387

☐ Laboratory ☒ On-Site

### Site Specifications

Yes No

**1 Secured Environment:**

Location must not be accessible to the general public.

☒

☐

**2 Environment free from contaminants**

☒

☐

**3 Dedicated telephone line:**

The phone line should be used solely by the instrument.

☒

☐

**4 Adequate Ventilation:**

Enough space must be provided for the instrument and the printer to allow for adequate ventilation.

☒

☐

**5 Stable temperature controls:**

The instrument should not be exposed to direct sources of temperature variation.

☒

☐

**6 115VAC power source:**

Whenever possible instrument power circuit should not be shared with refrigerators, copy machines or other high amperage devices.

☒

☐

**7 Room Condition:**

The site should be free of excessive moisture, heat and dust.

☒

☐

E-Seek	Old serial no.	New serial no.	Supplies
Keyboard	2000104859		Mouthpieces
Printer	7EHZ104724		Certificates of Analysis
UPS	7CECEB30502		Operator Worksheet
	JB0816006487		Other:

**Maintenance (if any)**

Firmware/Software Upgrade.

Upon installation at new location the phone cords were replaced.

**Comments (if any)**

Moved instrument from 128 Courthouse Lane, Clintwood VA, 24228 to 1019 Happy Valley Drive, Clintwood VA, 24228.

Instrument Serial Number 010606

Certification Date 9/13/2013

Instrument Barometer (mm HG) 706

Reference Barometer (mm HG) 709

Reference Barometer(RB)Serial # 009111

RB Calibration Due 5/23/2014

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.233
0.250	0.232	0.225	0.238	Sample 2	0.232
Precision		sample min	sample max	Sample 3	0.232
0.001		0.232	0.233		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.139
0.150	0.139	0.134	0.143	Sample 2	0.139
Precision		sample min	sample max	Sample 3	0.139
0		0.139	0.139		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.074
0.080	0.074	0.071	0.077	Sample 2	0.074
Precision		sample min	sample max	Sample 3	0.074
0		0.074	0.074		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.018
0.020	0.019	0.016	0.022	Sample 2	0.018
Precision		sample min	sample max	Sample 3	0.018
0		0.018	0.018		

all measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

☒ Calibrated☒ Certified☐ Replaced dry gas standard (+O-ring)

Dry gas standard Lot No. (with tank no.)

AG217702-20

☒ Measurement Assurance Check☒ Run a Breath Test (Instrument Test)

(successful completion denotes satisfactory condition of the item)

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.093
0.100	0.093	0.090	0.096	Sample 2	0.093
Precision		sample min	sample max	Sample 3	0.093
0		0.093	0.093		

DFS Technician

Date

9-19-13

Issuing Analyst

Date

9/22/13

Intox EC/IR II QA Worksheet  
Issued by Calibration and Training Program Manager  
Issue Date: 4-March-2013

DFS Document 250-F113  
Revision 6  
Page 2 of 2



COMMONWEALTH OF VIRGINIA  
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS  
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED <b>INSTRUMENT, TEST,</b>	NAME OF COURT <b>DFS</b>
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY <b>NEECE, JR., TIMMY, L</b>		AGENCY <b>DFS Central Lab</b>
DFS LICENSE NUMBER <b>25387</b>	LICENSE EXPIRES <b>10/01/2014</b>	DATE TEST CONDUCTED <b>09/13/2013</b>
TEST EQUIPMENT NUMBER <b>010606</b>		

RESULTS: TIME SAMPLE TAKEN 13:19 EDT  
SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS \_\_\_\_\_ DAY OF \_\_\_\_\_, 20\_\_\_\_\_.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS \_\_\_\_\_  
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS \_\_\_\_\_  
OPERATOR'S SIGNATURE



# IntoxNet MIS Report

Report Generated 13 Sep 2013 at 14:46

mmB  
TLN  
WHL

## Test Results

Instrument Serial Number 010606

Test # 000235 Subject Test

Test Location 1 Department of  
Test Date 13 Sep 2013

Test Location 2 Forensic Science  
Test Time 13:13  
Remote/Local Local

Test Location 3  
System Check Passed

Operator's Last Name NEECE, JR.

Agency DFS Central Lab

Card Serial Number 125387

Subject's Last Name INSTRUMENT

Subject's Middle Initial

Driver's License Number

Driver's License State

Operator's First Name TIMMY

Effective Date 10/01/2012

Subject's Date of Birth 00/00/0000

License Number 25387

Operator's Middle Initial L

Expiration Date 10/01/2014

Subject's First Name TEST

Subject's Sex Male

Driver's License Expiration 00/00/0000

Court Name DFS

End Date 13 Sep 2013 End Time 13:20

Result Time 13:19

Result Date 13 Sep 2013

Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 13:13

Data Type BLK

Sample Value 0.000

Sample Time 13:14

Data Type CHK

Sample Value 0.093

Sample Time 13:15

Data Type BLK

Sample Value 0.000

Sample Time 13:16

Data Type SUBJ

Sample Value 0.000

Sample Time 13:17

Data Type BLK

Sample Value 0.000

Sample Time 13:18

Data Type SUBJ

Sample Value 0.000

Sample Time 13:19

Data Type BLK

Sample Value 0.000

Sample Time 13:20

Standard Type Dry Gas Std

Standard Value 0.094

Standard Lot Number AG217702-20

Standard Expiration Date 06/25/2014

Tank Pressure 1046

Barometric Pressure 711 mmHg

Blow Sample Number 1 Blow Duration 3.50 sec

Blow Volume 1988 cc End-of-Blow Time 13:17

Blow Sample Number 2 Blow Duration 2.49 sec

Blow Volume 1688 cc End-of-Blow Time 13:19

Tamper Evident Stamp 916212da

Test Status Code 0

Test Status Success

TLN

# **Intox EC/IR-II: Accuracy Check**

NSK

*Department of Forensic Science*

Serial Number: 010606      Test Number: 236

Test Date: 09/13/2013      Test Time: 13:21 EDT

Dry Gas Target: 0.094

Lot Number: AG217702-20      Exp Date: 06/25/2014

Tank Pressure: 1041 psi      Barometric Pressure: 711 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	13:22
CHK	0.092	13:23
BLK	0.000	13:24
CHK	0.092	13:25
BLK	0.000	13:26
CHK	0.092	13:27

Test Status: Success

Calibration CRC: 36044A1A

**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010606      Test Number: 237

Test Date: 09/13/2013      Test Time: 13:31 EDT

Dry Gas Target: 0.281

Lot Number: AG322702-03      Exp Date: 08/15/2015

Tank Pressure: 477 psi      Barometric Pressure: 711 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	13:32
CHK	0.277	13:33
BLK	0.000	13:34
CHK	0.277	13:35
BLK	0.000	13:36
CHK	0.276	13:37

Test Status: *Success*

Calibration CRC: 36044A1A

TLN  
N8K





COMMONWEALTH OF VIRGINIA  
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS  
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED <b>INSTRUMENT, TEST,</b>	NAME OF COURT <b>DFS</b>
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY <b>NEECE, JR., TIMMY, L</b>		AGENCY <b>DFS Central Lab</b>
DFS LICENSE NUMBER <b>25387</b>	LICENSE EXPIRES <b>10/01/2014</b>	DATE TEST CONDUCTED <b>09/13/2013</b>
TEST EQUIPMENT NUMBER <b>010606</b>		

RESULTS: TIME SAMPLE TAKEN 14:36 EDT  
SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS \_\_\_\_\_ DAY OF \_\_\_\_\_, 20\_\_\_\_.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS \_\_\_\_\_  
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS \_\_\_\_\_  
OPERATOR'S SIGNATURE



# IntoxNet MIS Report

Report Generated 13 Sep 2013 at 14:45

mmB  
TLN  
WJL

## Test Results

Instrument Serial Number 010606

Test # 000238

Subject Test

Test Location 1 Department of

Test Date 13 Sep 2013

Test Location 2 Forensic Science

Test Time 14:30

Remote/Local Local

Test Location 3

System Check Passed

Operator's Last Name NEECE, JR.

Agency DFS Central Lab

Card Serial Number 125387

Subject's Last Name INSTRUMENT

Subject's Middle Initial

Driver's License Number

Driver's License State

End Date 13 Sep 2013

End Time 14:37

Data Type DIAG

Data Type BLK

Data Type CHK

Data Type BLK

Data Type SUBJ

Data Type BLK

Data Type SUBJ

Data Type BLK

Standard Type Dry Gas Std

Standard Lot Number AG217702-20

Tank Pressure 1012

Blow Sample Number 1

Blow Duration 2.76 sec

Blow Sample Number 2

Blow Duration 3.15 sec

Tamper Evident Stamp d6d4b47f

Test Status Success

Operator's First Name TIMMY

Effective Date 10/01/2012

Subject's Date of Birth 00/00/0000

License Number 25387

Operator's Middle Initial L

Expiration Date 10/01/2014

Subject's First Name TEST

Subject's Sex Male

Driver's License Expiration 00/00/0000

Court Name DFS

Result Time 14:36

Result Date 13 Sep 2013

Result 0.00

Sample Value Pass

Sample Value 0.000

Sample Value 0.091

Sample Value 0.000

Sample Value 0.000

Sample Value 0.000

Sample Value 0.000

Sample Value 0.000

Sample Time 14:30

Sample Time 14:31

Sample Time 14:32

Sample Time 14:33

Sample Time 14:34

Sample Time 14:34

Sample Time 14:36

Sample Time 14:37

Standard Value 0.093

Standard Expiration Date 06/25/2014

Barometric Pressure 706 mmHg

Blow Volume 1777 cc

End-of-Blow Time 14:34

Blow Volume 1966 cc

End-of-Blow Time 14:36

Test Status Code 0

**Intox EC/IR-II: Software Version**

*Department of Forensic Science*

Serial Number: 010606

System Date: 09/13/2013

System Time: 15:04 EDT

Software Version: VS00333-K

Software CRC: 56745BCA

Firmware Version: VS00171-CL

Firmware CRC: 7107756B

TLN

Net

**Intox EC/IR-II: Calibration**

*Department of Forensic Science*

Serial Number: 010606      Test Number: 239

Test Date: 09/13/2013      Test Time: 15:08 EDT

Dry Gas Target: 0.093

Lot Number: 0651210013-01      Exp Date: 04/01/2014

Tank Pressure: 865 psi      Barometric Pressure: 705 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	15:09
CAL	0.093	15:09

Test Status: *Success*

Calibration CRC: 35EC39BD

TLN

NRK

**Intox EC/IR-II: Certification**

TLN

**Department of Forensic Science**

Serial Number: 010606      Test Number: 240  
Test Date: 09/13/2013      Test Time: 15:13 EDT

WJL

Technician Initials: TLN

Dry Gas Target: 0.232

Lot Number: 0651225015-02      Exp Date: 04/01/2014

Tank Pressure: 708 psi      Barometric Pressure: 705 mmHg

System Check: Passed

Test	g/210L	Time
DIAG	Pass	15:13
BLK	0.000	15:14
CHK	0.233	15:15
BLK	0.000	15:16
CHK	0.232	15:17
BLK	0.000	15:18
CHK	0.232	15:19
BLK	0.000	15:21

**Success**

Standard Value (at sea level): 0.250 g/210L

Certification Status: Continued

**Intox EC/IR-II: Certification**

*Department of Forensic Science*

Serial Number: 010606      Test Number: 241  
Test Date: 09/13/2013      Test Time: 15:23 EDT

Technician Initials: TLN

Dry Gas Target: 0.139

Lot Number: 0651215014-06      Exp Date: 04/01/2014

Tank Pressure: 169 psi      Barometric Pressure: 705 mmHg

System Check: Passed

Test	g/210L	Time
DIAG	Pass	15:24
BLK	0.000	15:25
CHK	0.139	15:25
BLK	0.000	15:27
CHK	0.139	15:27
BLK	0.000	15:29
CHK	0.139	15:30
BLK	0.000	15:31

**Success**

Standard Value (at sea level): 0.150 g/210L

Certification Status: Continued

**Intox EC/IR-II: Certification**

*Department of Forensic Science*

Serial Number: 010606      Test Number: 242  
Test Date: 09/13/2013      Test Time: 15:34 EDT

Technician Initials: TLN

Dry Gas Target: 0.074

Lot Number: 0651208012-06      Exp Date: 04/01/2014

Tank Pressure: 213 psi      Barometric Pressure: 705 mmHg

System Check: Passed

Test	g/210L	Time
DIAG	Pass	15:34
BLK	0.000	15:35
CHK	0.074	15:36
BLK	0.000	15:37
CHK	0.074	15:38
BLK	0.000	15:39
CHK	0.074	15:40
BLK	0.000	15:41

**Success**

Standard Value (at sea level): 0.080 g/210L

Certification Status: Continued

TLN

AKL

**Intox EC/IR-II: Certification**

*Department of Forensic Science*

Serial Number: 010606      Test Number: 243  
Test Date: 09/13/2013      Test Time: 15:44 EDT

Technician Initials: TLN

Dry Gas Target: 0.019

Lot Number: 0651202011-07      Exp Date: 04/01/2014

Tank Pressure: 301 psi      Barometric Pressure: 705 mmHg

System Check: Passed

Test	g/210L	Time
DIAG	Pass	15:44
BLK	0.000	15:45
CHK	0.018	15:45
BLK	0.000	15:47
CHK	0.018	15:47
BLK	0.000	15:49
CHK	0.018	15:50
BLK	0.000	15:51

**Success**

Standard Value (at sea level): 0.020 g/210L

Certification Status: Success



**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010606      Test Number: 244

Test Date: 09/13/2013      Test Time: 15:55 EDT

Dry Gas Target: 0.278

Lot Number: AG322702-03      Exp Date: 08/15/2015

Tank Pressure: 458 psi      Barometric Pressure: 705 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	15:56
CHK	0.280	15:56
BLK	0.000	15:58
CHK	0.281	15:59
BLK	0.000	16:01
CHK	0.280	16:02

Test Status: *Success*

Calibration CRC: 35EC39BD

TLN

NJL

**Intox EC/IR-II: Accuracy Check**

*Department of Forensic Science*

Serial Number: 010606      Test Number: 245

Test Date: 09/13/2013      Test Time: 16:05 EDT

Dry Gas Target: 0.093

Lot Number: AG217702-20      Exp Date: 06/25/2014

Tank Pressure: 992 psi      Barometric Pressure: 705 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	16:06
CHK	0.093	16:07
BLK	0.000	16:08
CHK	0.093	16:09
BLK	0.000	16:10
CHK	0.093	16:11

Test Status: Success

Calibration CRC: 35EC39BD



COMMONWEALTH OF VIRGINIA  
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS  
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED <b>INSTRUMENT, TEST,</b>	NAME OF COURT <b>DFS</b>
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY <b>NEECE, JR., TIMMY, L</b>		AGENCY <b>DFS Central Lab</b>
DFS LICENSE NUMBER <b>25387</b>	LICENSE EXPIRES <b>10/01/2014</b>	DATE TEST CONDUCTED <b>09/13/2013</b>
TEST EQUIPMENT NUMBER <b>010606</b>		

RESULTS: TIME SAMPLE TAKEN 16:20 EDT  
SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS \_\_\_\_\_ DAY OF \_\_\_\_\_, 20\_\_\_\_.

BREATH TEST OPERATOR
----------------------

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS \_\_\_\_\_  
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS \_\_\_\_\_  
OPERATOR'S SIGNATURE



# IntoxNet MIS Report

Report Generated 19 Sep 2013 at 09:06

TLN  
mll

## Test Results

Instrument Serial Number 010606

Test # 000246 Subject Test

Test Location 1 Department of  
Test Date 13 Sep 2013

Test Location 2 Forensic Science  
Test Time 16:14  
Remote/Local Local

Test Location 3  
System Check Passed

Operator's Last Name NEECE, JR.  
Agency DFS Central Lab

Operator's First Name TIMMY

Operator's Middle Initial L

Card Serial Number 125387

License Number 25387

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 13 Sep 2013 End Time 16:21

Result Time 16:20

Result Date 13 Sep 2013 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 16:14

Data Type BLK

Sample Value 0.000

Sample Time 16:15

Data Type CHK

Sample Value 0.093

Sample Time 16:16

Data Type BLK

Sample Value 0.000

Sample Time 16:17

Data Type SUBJ

Sample Value 0.000

Sample Time 16:18

Data Type BLK

Sample Value 0.000

Sample Time 16:19

Data Type SUBJ

Sample Value 0.000

Sample Time 16:20

Data Type BLK

Sample Value 0.000

Sample Time 16:21

Standard Type Dry Gas Std

Standard Value 0.093

Standard Lot Number AG217702-20

Standard Expiration Date 06/25/2014

Tank Pressure 982

Barometric Pressure 705 mmHg

Blow Sample Number 1 Blow Duration 2.74 sec

Blow Volume 1826 cc End-of-Blow Time 16:18

Blow Sample Number 2 Blow Duration 2.79 sec

Blow Volume 2077 cc End-of-Blow Time 16:20

Tamper Evident Stamp 71d9b471

Test Status Code 0

Test Status Success